

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3569

Client ID : SSTDCCC040

Date Analyzed: 11/13/2025

Lab File ID: BF144243.D

Time Analyzed: 14:56

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	57801	6.869	225618	8.15	126474	9.90
UPPER LIMIT	115602	7.369	451236	8.651	252948	10.404
LOWER LIMIT	28900.5	6.369	112809	7.651	63237	9.404
EPA SAMPLE NO.						
01 PB170473BL	58185	6.86	226377	8.15	128860	9.90
02 PB170473BS	61443	6.87	243453	8.15	140132	9.91
03 PB170473BSD	58832	6.86	241570	8.15	135120	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.