

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3569

Client ID : SSTDCCC040

Date Analyzed: 11/12/2025

Lab File ID: BP026108.D

Time Analyzed: 11:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	274793	7.672	1101110	10.44	695884	14.32
UPPER LIMIT	549586	8.172	2202220	10.937	1391770	14.819
LOWER LIMIT	137397	7.172	550555	9.937	347942	13.819
EPA SAMPLE NO.						
01 MW-12DL	345995	7.67	1363080	10.44	867470	14.31

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.