

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3569

Client ID: SSTDCCC040

Date Analyzed: 11/12/2025

Lab File ID: BP026108.D

Time Analyzed: 11:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1413050	17.125	1582920	21.583	1849640	24.924
UPPER LIMIT	2826100	17.625	3165840	22.083	3699280	25.424
LOWER LIMIT	706525	16.625	791460	21.083	924820	24.424
EPA SAMPLE NO.						
MW-12DL	1776020	17.13	2024260	21.58	2316390	24.90

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.