

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP026027.D
Instrument ID: BNA_P

Contract: REMI02
SDG NO.: Q3569
DFTPP Injection Date: 10/29/2025
DFTPP Injection Time: 08:45

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	32.9
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	11.5
442	Greater than 50% of mass 198	75
443	15.0 - 24.0% of mass 442	14.4 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP026028.D	10/29/2025	09:26
SSTDICC005	SSTDICC005	BP026029.D	10/29/2025	10:07
SSTDICC010	SSTDICC010	BP026030.D	10/29/2025	10:48
SSTDICC020	SSTDICC020	BP026031.D	10/29/2025	11:29
SSTDICCC040	SSTDICCC040	BP026032.D	10/29/2025	12:10
SSTDICC050	SSTDICC050	BP026033.D	10/29/2025	12:51
SSTDICC060	SSTDICC060	BP026034.D	10/29/2025	13:32
SSTDICC080	SSTDICC080	BP026035.D	10/29/2025	14:13