

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
Lab Code: ACE SDG NO.: Q3569
Lab File ID: VX048527.D Date Analyzed: 11/10/2025
Instrument ID: MSVOA_X Time Analyzed: 09:36
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	178330	5.53	288528	6.74	248957	10.04
UPPER LIMIT	356660	6.032	577056	7.239	497914	10.537
LOWER LIMIT	89165	5.032	144264	6.239	124479	9.537
EPA SAMPLE NO.						
MW-2	226787	5.54	431937	6.75	431296	10.04
MW-12	174503	5.54	338583	6.75	347291	10.04
TB-2	185085	5.54	364726	6.75	365529	10.04
VX1110WBL01	157554	5.54	305452	6.74	313169	10.04
VX1110WBS01	140876	5.54	243139	6.75	219757	10.04
VX1110WBSD01	186739	5.53	314084	6.75	272722	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.