

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
 Lab Code: ACE SDG NO.: Q3569
 Lab File ID: VX048549.D Date Analyzed: 11/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:11
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	181006	5.53	289835	6.74	252348	10.04
UPPER LIMIT	362012	6.032	579670	7.239	504696	10.537
LOWER LIMIT	90503	5.032	144918	6.239	126174	9.537
EPA SAMPLE NO.						
MW-12DL	220460	5.54	428617	6.75	439851	10.04
VX1111WBL01	184346	5.54	353648	6.75	374656	10.04
VX1111WBS01	187478	5.53	320628	6.74	287293	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.