

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02  
 Lab Code: ACE SDG NO.: Q3569  
 Lab File ID: VX048527.D Date Analyzed: 11/10/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:36  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 127604        | 12.00€ |  |  |  |  |
| UPPER LIMIT    | 255208        | 12.50€ |  |  |  |  |
| LOWER LIMIT    | 63802         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| MW-2           | 228614        | 12.01  |  |  |  |  |
| MW-12          | 188428        | 12.01  |  |  |  |  |
| TB-2           | 198620        | 12.01  |  |  |  |  |
| VX1110WBL01    | 160725        | 12.01  |  |  |  |  |
| VX1110WBS01    | 112690        | 12.01  |  |  |  |  |
| VX1110WBSD01   | 139918        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.