

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3569

Lab File ID: VX048548.D

BFB Injection Date: 11/11/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:09

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.5
75	30.0 - 60.0% of mass 95	48.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	76.9
175	5.0 - 9.0% of mass 174	5.7 (7.4) 1
176	95.0 - 101.0% of mass 174	76.1 (99) 1
177	5.0 - 9.0% of mass 176	5 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048549.D	11/11/2025	09:11
VX1111WBL01	VX1111WBL01	VX048551.D	11/11/2025	10:02
VX1111WBS01	VX1111WBS01	VX048552.D	11/11/2025	10:35
MW-12DL	Q3569-02DL	VX048557.D	11/11/2025	12:24