

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3573
 Lab File ID: VX048527.D Date Analyzed: 11/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:36
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	178330	5.53	288528	6.74	248957	10.04
UPPER LIMIT	356660	6.032	577056	7.239	497914	10.537
LOWER LIMIT	89165	5.032	144264	6.239	124479	9.537
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	154204	5.54	305047	6.75	320823	10.04
STORAGE-BLANK-WATER-REF-5	177862	5.54	361031	6.75	373583	10.04
STORAGE-BLANK-WATER-REF-4	159408	5.54	314204	6.75	266311	10.04
STORAGE-BLANK-SAMPLE-MANAGEMENT	155276	5.54	330639	6.75	321826	10.04
VX1110WBL01	157554	5.54	305452	6.74	313169	10.04
VX1110WBS01	140876	5.54	243139	6.75	219757	10.04
VX1110WBSD01	186739	5.53	314084	6.75	272722	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.