

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3573
 Lab File ID: VX048527.D Date Analyzed: 11/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:36
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	127604	12.00€				
UPPER LIMIT	255208	12.50€				
LOWER LIMIT	63802	11.50€				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	164496	12.01				
STORAGE-BLANK-WATER-REF-5	199089	12.01				
STORAGE-BLANK-WATER-REF-4	138516	12.01				
STORAGE-BLANK-SAMPLE-MANAGEMENT	147785	12.01				
VX1110WBL01	160725	12.01				
VX1110WBS01	112690	12.01				
VX1110WBSD01	139918	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.