

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137843

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/10/2025	13:33
2	CAL2	1	Water	NA	NA	20.2	7.00	11/10/2025	13:35
3	CAL3	1	Water	NA	NA	20.3	10.02	11/10/2025	13:36
4	ICV	1	Water	NA	NA	20.3	7.02	11/10/2025	13:37
5	CCV1	1	Water	NA	NA	20.2	2.01	11/10/2025	13:39
6	Q3569-01	1	Water	NA	NA	20.5	6.88	11/10/2025	13:44
7	Q3569-02	1	Water	NA	NA	20.7	6.49	11/10/2025	13:47
8	Q3574-01	1	Water	NA	NA	20.2	6.63	11/10/2025	13:55
9	Q3578-01	1	Water	NA	NA	20.2	5.02	11/10/2025	14:00
10	Q3584-01	1	Water	NA	NA	20.6	6.73	11/10/2025	14:10
11	Q3584-02	1	Water	NA	NA	20.5	6.39	11/10/2025	14:15
12	Q3584-03	1	Water	NA	NA	20.5	6.60	11/10/2025	14:22
13	Q3584-04	1	Water	NA	NA	20.9	7.40	11/10/2025	14:25
14	Q3584-05	1	Water	NA	NA	20.5	7.30	11/10/2025	14:30
15	Q3584-07	1	Water	NA	NA	20.2	7.60	11/10/2025	14:37
16	CCV2	1	Water	NA	NA	20.2	12.02	11/10/2025	14:40
17	Q3595-01	1	Water	NA	NA	22.8	7.13	11/10/2025	14:48
18	Q3595-01DUP	1	Water	NA	NA	22.9	7.15	11/10/2025	14:50
19	CCV3	1	Water	NA	NA	20.3	2.01	11/10/2025	15:00

171843

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH Q3584 WorkList ID : 193013 Department : Wet-Chemistry Date : 11-10-2025 12:56:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3569-01	MW-2	Water	pH	Cool 4 deg C	REMI02	J22	11/05/2025	9040C
Q3569-02	MW-12	Water	pH	Cool 4 deg C	REMI02	J22	11/05/2025	9040C
Q3574-01	304641-01 Liquid	Water	pH	Cool 4 deg C	PACI01	J22	11/04/2025	9040C
Q3578-01	MH-1172025	Water	pH	Cool 4 deg C	EURO03	D31	11/07/2025	9040C
Q3584-01	SW-1	Water	pH	Cool 4 deg C	REMI02	D41	11/06/2025	9040C
Q3584-02	SW-2	Water	pH	Cool 4 deg C	REMI02	D41	11/07/2025	9040C
Q3584-03	SW-3	Water	pH	Cool 4 deg C	REMI02	D41	11/07/2025	9040C
Q3584-04	EB-2	Water	pH	Cool 4 deg C	REMI02	D41	11/06/2025	9040C
Q3584-05	FB-2	Water	pH	Cool 4 deg C	REMI02	D41	11/07/2025	9040C
Q3584-07	SEEP-1	Water	pH	Cool 4 deg C	REMI02	D41	11/07/2025	9040C
Q3595-01	LAW-25-0176	Water	pH	Cool 4 deg C	PSEG03	D41	11/07/2025	9040C

Date/Time 11-10-25 13:30
Raw Sample Received by: SB
Raw Sample Relinquished by: CS

Date/Time 11-10-25 14:30
Raw Sample Received by: CS
Raw Sample Relinquished by: SB