

Initial and Continuing Calibration Verification

Client: Pacific Commercial Services Inc.

SDG No.: Q3574

Project: Kilo Pier

RunNo.: LB137843

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.02	7	100	90-110	11/10/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	11/10/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	11/10/2025
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	11/10/2025

Initial and Continuing Calibration Verification

Client: Pacific Commercial Services Inc.

SDG No.: Q3574

Project: Kilo Pier

RunNo.: LB137966

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Flash Point		o F	82.4	81	102	78-84	11/19/2025