

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Pacific Commercial Services Inc.
Contract: PACI01
ICS Source: EPA

SDG No.: Q3574
Lab Code: ACE
Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	7.40			-20	20	11/12/2025	17:44	LB137883
	Barium	8.38	6.0	140	-94	106	11/12/2025	17:44	LB137883
	Cadmium	-0.82	1.0	82	-5	7	11/12/2025	17:44	LB137883
	Chromium	50.4	52.0	97	42	62	11/12/2025	17:44	LB137883
	Lead	-8.49			-12	12	11/12/2025	17:44	LB137883
	Selenium	-3.77			-20	20	11/12/2025	17:44	LB137883
	Silver	8.44			-10	10	11/12/2025	17:44	LB137883
ICSAB01	Arsenic	113	104	109	88.4	120	11/12/2025	17:48	LB137883
	Barium	464	537	86	437	637	11/12/2025	17:48	LB137883
	Cadmium	966	972	99	826	1120	11/12/2025	17:48	LB137883
	Chromium	558	542	103	460	624	11/12/2025	17:48	LB137883
	Lead	38.8	49.0	79	37	61	11/12/2025	17:48	LB137883
	Selenium	38.7	46.0	84	26	66	11/12/2025	17:48	LB137883
	Silver	204	201	102	170	232	11/12/2025	17:48	LB137883