

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Pacific Commercial Services Inc.

SDG No.: Q3574

Contract: PACI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170507BS							
Arsenic	ug/L	400	388		97	87 - 113	P
Barium	ug/L	100	95.5		96	88 - 113	P
Cadmium	ug/L	100	91.0		91	88 - 113	P
Chromium	ug/L	200	197		98	90 - 113	P
Lead	ug/L	500	447		89	86 - 113	P
Selenium	ug/L	1000	928		93	83 - 114	P
Silver	ug/L	37.5	36.1		96	84 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Pacific Commercial Services Inc.

SDG No.: Q3574

Contract: PACI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170511BS Mercury	ug/L	4.0	3.90		98	82 - 119	CV