

## Report of Analysis

Client:	Pacific Commercial Services Inc.	Date Collected:	
Project:	Kilo Pier	Date Received:	
Client Sample ID:	BSF1110W2	SDG No.:	Q3574
Lab Sample ID:	BSF1110W2	Matrix:	Water
Analytical Method:	8015D GRO	% Solid:	0
Sample Wt/Vol:	5 mL	Test:	Gasoline Range Organics
	Final Vol: 5 mL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
<b>TARGETS</b>										
GRO	GRO	198		1	6.00	9.00	45.0	ug/L	11/10/25 10:09	FB111025
<b>SURROGATES</b>										
98-08-8	Alpha,Alpha,Alpha-Trifluor	20.7			50 - 150		103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit