

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: PACI01

ProjectID: Kilo Pier

Lab Code: ACE

SDG No.: Q3574

Calibration Sequence : FF102925		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
85	9472915	111446	FF016603.D
170	18165984	106859	FF016604.D
340	37023074	108891	FF016605.D
850	89698209	105527	FF016606.D
1700	169764046	99861	FF016607.D
AVG RF : 106517		% RSD : 4.074	AVG RT : 15.059