

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: PACI01
ProjectID: Kilo Pier
Lab Code: ACE SDG No.: Q3574
DataFile: FF016645.D Analyst Name: YP\AJ Analyst Date: 11-11-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	85848516	100998	106517	5.181