

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3574

Analytical Method: 8270E

Client: Pacific Commercial Services Inc.

DataFile: BF144245.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170473BS	Naphthalene	50	41.6	ug/L	83				40	121	
	Acenaphthylene	50	46.9	ug/L	94				41	130	
	Acenaphthene	50	37.6	ug/L	75				47	122	
	Fluorene	50	42.8	ug/L	86				52	124	
	Phenanthrene	50	43.2	ug/L	86				59	120	
	Anthracene	50	43.3	ug/L	87				57	123	
	Fluoranthene	50	41.3	ug/L	83				57	128	
	Pyrene	50	51.7	ug/L	103				57	126	
	Benzo(a)anthracene	50	45.2	ug/L	90				58	125	
	Chrysene	50	46.5	ug/L	93				59	123	
	Benzo(b)fluoranthene	50	41.5	ug/L	83				53	131	
	Benzo(k)fluoranthene	50	46.1	ug/L	92				57	129	
	Benzo(a)pyrene	50	46.7	ug/L	93				54	128	
	Indeno(1,2,3-cd)pyrene	50	45.9	ug/L	92				52	134	
	Dibenz(a,h)anthracene	50	45.7	ug/L	91				51	134	
	Benzo(g,h,i)perylene	50	47.4	ug/L	95				50	134	