

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PACI01
Lab Code: ACE SDG NO.: Q3574
Lab File ID: VX048563.D Date Analyzed: 11/12/2025
Instrument ID: MSVOA_X Time Analyzed: 09:11
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	168053	5.53	278642	6.74	246918	10.04
UPPER LIMIT	336106	6.025	557284	7.239	493836	10.537
LOWER LIMIT	84026.5	5.025	139321	6.239	123459	9.537
EPA SAMPLE NO.						
304641-01 Liquid	175785	5.54	342013	6.75	348551	10.04
VX1112WBL01	211958	5.54	425634	6.75	439104	10.04
VX1112WBS01	191720	5.53	311333	6.74	276234	10.04
VX1112WBSD01	161754	5.53	273225	6.74	245374	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.