

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3575

Project: Transfer Station-SPDES

Run No.: LB137848

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137848BS							
TSS	mg/L	550	531		96	1	90-110	11/11/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3575

Project: Transfer Station-SPDES

Run No.: LB137871

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137871BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	11/07/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3575

Project: Transfer Station-SPDES

Run No.: LB137877

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137877BS							
Oil and Grease	mg/L	20.0	18.2		91	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3575

Project: Transfer Station-SPDES

Run No.: LB137852

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170466BS							
Ammonia as N	mg/L	1	0.97		97	1	90-110	11/11/2025