

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q3575

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	18.7	2.0	935	-18	22	11/10/2025	12:21	LB137841
	Iron	103000	101000	102	85600	116500	11/10/2025	12:21	LB137841
	Lead	-2.05			-12	12	11/10/2025	12:21	LB137841
ICSAB01	Copper	501	511	98	434	588	11/10/2025	12:26	LB137841
	Iron	103000	99300	104	84400	114500	11/10/2025	12:26	LB137841
	Lead	49.4	49.0	101	37	61	11/10/2025	12:26	LB137841
ICSA	Copper	15.6	2.0	780	-18	22	11/10/2025	12:30	LB137841
	Iron	92800	101000	92	85600	116500	11/10/2025	12:30	LB137841
	Lead	-17.2			-12	12	11/10/2025	12:30	LB137841
ICSAB	Copper	490	511	96	434	588	11/10/2025	12:34	LB137841
	Iron	91900	99300	92	84400	114500	11/10/2025	12:34	LB137841
	Lead	7.69	49.0	16	37	61	11/10/2025	12:34	LB137841