



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3575 Analytical Method: SW624.1  
Client: Tully Environmental, Inc Datafile : VN088283.D

| Lab Sample ID | Parameter     | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits | RPD |
|---------------|---------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |               |       |        |      |     |     |      |     | High   |     |
| VN1114WBS01   | Benzene       | 20    | 19.4   | ug/L | 97  |     |      | 65  | 135    |     |
|               | Toluene       | 20    | 19.7   | ug/L | 99  |     |      | 70  | 130    |     |
|               | Ethyl Benzene | 20    | 19.2   | ug/L | 96  |     |      | 60  | 140    |     |
|               | m/p-Xylenes   | 40    | 38.7   | ug/L | 97  |     |      | 87  | 111    |     |
|               | o-Xylene      | 20    | 18.9   | ug/L | 95  |     |      | 87  | 111    |     |