

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
Lab Code: ACE SDG NO.: Q3575
Lab File ID: VN088282.D Date Analyzed: 11/14/2025
Instrument ID: MSVOA_N Time Analyzed: 09:07
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	56753	7.79	292349	9.08	258679	11.85
UPPER LIMIT	113506	8.294	584698	9.582	517358	12.347
LOWER LIMIT	28376.5	7.294	146175	8.582	129340	11.347
EPA SAMPLE NO.						
001 Willets Pt Blvd (Nov)	54259	7.79	301359	9.08	265781	11.85
002 35th Ave (Nov)	55641	7.80	290103	9.08	257187	11.84
VN1114WBL01	53945	7.79	285107	9.08	248105	11.84
VN1114WBS01	59304	7.79	313554	9.08	280371	11.84
VN1114WBSD01	57783	7.79	307983	9.08	272076	11.85

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.