

DATA PACKAGE

GENERAL CHEMISTRY

PROJECT NAME : MCUA PERMIT NO 14241 - 571 JERSEY AVE NB NJ

EUROPASTRY

571 Jersey Ave

New Brunswick, NJ - 08901

Phone No: 631-563-6262 x2602

ORDER ID : Q3578

ATTENTION : Kevin Carlucci



Laboratory Certification ID # 20012



Table Of Contents for Q3578

1) Signature Page	3
2) Case Narrative	5
3) Qualifier Page	7
4) QA Checklist	8
5) Genchem Data	9
6) Shipping Document	33
6.1) CHAIN OF CUSTODY	34
6.2) Lab Certificate	37

1
2
3
4
5
6

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

1

Laboratory Name : Alliance Technical Group LLC Client : Europastry
 Project Location : NJ Project Number : MCUA Permit No 14241 - 571 Jersey Ave NB NJ
 Laboratory Sample ID(s) : Q3578 Sampling Date(s) : 11/07/2025
 List DKQP Methods Used (e.g., 8260,8270, et Cetra) **1664A,9040C,Sampling,SM2540 D,SM5210 B,SM5220 D,SOP**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
1A	Were the method specified handling, preservation, and holding time requirements met?	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
3	Were samples received at an appropriate temperature (4±2° C)?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

Cover Page

Order ID : Q3578

Project ID : MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Client : Europastry

Lab Sample Number

Q3578-01
Q3578-02
Q3578-03

Client Sample Number

MH-1172025
MH-1172025MS
MH-1172025MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:08 pm, Nov 18, 2025

Date: 11/14/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Europastry

Project Name: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Project # N/A

Order ID # Q3578

Test Name: BOD5,COD,Oil and Grease,pH,TPH,TSS

A. Number of Samples and Date of Receipt:

3 Water samples were received on 11/07/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested:

BOD5,COD,Oil and Grease,pH,TPH,TSS. This data package contains results for

BOD5,COD,Oil and Grease,pH,TPH,TSS.

C. Analytical Techniques:

The analysis of Oil and Grease,TPH was based on method 1664A, The analysis of pH was based on method 9040C, The analysis of TSS was based on method SM2540 D, The analysis of BOD5 was based on method SM5210 B and The analysis of COD was based on method SM5220 D.

D. QA/ QC Samples:

The Holding Times were met for all samples except for MH-1172025 of pH as sample was receive out of holding time.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Sample Q3578-01 was analyzed straight Dilution for COD Parameter because the original samples were reading over range, only 10X has been reported.

As per method 1664A(TPH), MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:08 pm, Nov 18, 2025

Signature _____

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3578

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 11/14/2025

LAB CHRONICLE

OrderID:	Q3578	OrderDate:	11/7/2025 1:38:00 PM
Client:	Europastray	Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ
Contact:	Kevin Carlucci	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3578-01	MH-1172025	WATER			11/07/25 08:50			11/07/25
			BOD5	SM5210 B			11/07/25 19:00	
			COD	SM5220 D			11/14/25 14:07	
			Oil and Grease	1664A			11/13/25 10:00	
			pH	9040C			11/10/25 14:00	
			TPH	1664A			11/13/25 10:00	
			TSS	SM2540 D			11/11/25 09:40	



SAMPLE DATA

Report of Analysis

Client:	Europastry	Date Collected:	11/07/25 08:50
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Date Received:	11/07/25
Client Sample ID:	MH-1172025	SDG No.:	Q3578
Lab Sample ID:	Q3578-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
BOD5	469		1	0.20	2.00	mg/L		11/07/25 19:00	SM 5210 B-16
COD	866	D	10	15.0	100	mg/L		11/14/25 14:07	SM 5220 D-11
Oil and Grease	94.2		1	0.29	5.00	mg/L		11/13/25 10:00	1664A
pH	5.02	H	1	0	0	pH		11/10/25 14:00	9040C
TPH	32.2		1	0.29	5.00	mg/L		11/13/25 10:00	1664A
TSS	59.9		1	1.00	4.00	mg/L		11/11/25 09:40	SM 2540 D-20

Comments: pH result reported at temperature 20.2 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3578

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137843

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.02	7	100	90-110	11/10/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	11/10/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	11/10/2025
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	11/10/2025

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3578

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137904

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	11/14/2025
Sample ID: CCV2 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV3 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV4 COD	mg/L	49.161	50	98	95-105	11/14/2025

Initial and Continuing Calibration Blank Summary

Client: Europastry

SDG No.: Q3578

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137904

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	1.50	10	09/08/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025
Sample ID: CCB4 COD	mg/L	< 5.0000	5.0000	U	1.50	10	11/14/2025

Preparation Blank Summary

Client: Europastry

SDG No.: Q3578

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137848BL TSS	mg/L	< 2.0000	2.0000	U	1	4	11/11/2025
Sample ID: LB137871BL BOD5	mg/L	< 0.2000	0.2000	U	0.20	2.0	11/07/2025
Sample ID: LB137877BL Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	11/13/2025
Sample ID: LB137878BL TPH	mg/L	< 2.5000	2.5000	U	0.29	5.0	11/13/2025
Sample ID: LB137904BL COD	mg/L	< 5.0000	5.0000	U	1.5	10.0	11/14/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3578-01
Client ID:	MH-1172025MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	115		94.2		20.0	1	102		11/13/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3578-01
Client ID:	MH-1172025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	113		94.2		20.0	1	95		11/13/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3630-05
Client ID:	DSN003MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	68.4		23.9		50.0	1	89		11/14/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3630-05
Client ID:	DSN003MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	67.4		23.9		50.0	1	87		11/14/2025

Duplicate Sample Summary

Client: Europastry Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ Client ID: LB137878BSD	SDG No.: Q3578 Sample ID: LB137878BS Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	19.0		17.8		1	6.52		11/13/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3578-01
Client ID:	MH-1172025DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
BOD5	mg/L	+/-20	469		466		1	0.64		11/07/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3578-01
Client ID:	MH-1172025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	115		113		1	1.23		11/13/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3595-01
Client ID:	LAW-25-0176DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
pH	pH	+/-20	7.13		7.15		1	0.28		11/10/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3630-05
Client ID:	DSN003DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
COD	mg/L	+/-20	23.9		22.9		1	4.27		11/14/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3630-05
Client ID:	DSN003MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
COD	mg/L	+/-20	68.4		67.4		1	1.47		11/14/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137848

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137848BS							
TSS	mg/L	550	531		96	1	90-110	11/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137871

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137871BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	11/07/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137877

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137877BS							
Oil and Grease	mg/L	20.0	18.2		91	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BS							
TPH	mg/L	20.0	19.0		95	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BSD							
TPH	mg/L	20.0	17.8		89	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137904

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137904BS							
COD	mg/L	50	53.2		106	1	90-110	11/14/2025



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: **EUROPASTAL**
ADDRESS: **571 JERSEY AVE**
CITY **NEW BRUNSWICK** STATE: **NJ** ZIP: **08901**
ATTENTION: **KEVIN CARLUCCI**
PHONE: **631-563-6262** FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **MCUA Permit No. 14241-571 Jersey**
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data) ☐ Other
☐ EDD FORMAT

PH BOD5 TSS COD ORG, TPH
1 2 3 4 5 6 7 8 9

ALLIANCE
SAMPLE
ID

PROJECT
SAMPLE IDENTIFICATION

SAMPLE
MATRIX

SAMPLE
TYPE
COMP GRAB

SAMPLE
COLLECTION
DATE TIME

OF BOTTLES

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

1.	MH-1172025	W	X	11-7	0850	8	X	X	X	X	X							MS/MSD
2.																		
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1.	DATE/TIME: 0905 11-7-25	RECEIVED BY: 1.	Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP. 4.9°C Comments: *PH. 5.32 TEMP. 23.3°C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	
RELINQUISHED BY SAMPLER: 3.	DATE/TIME: 1005 11-7-25	RECEIVED BY: 3.	

Page **1** of **1**

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Client Name: EUROPASTALY
Client Address: 57 JENSEY AVE
Client Rep on Site: KEVIN CALIRCI
Sampling Date: 11-4-25
Arrival Time: 0830

Project Name: MCUA Permit No. 14241-57 JENSEY AVE
Project Location: NEW BRUNSWICK
Cooler Custody Seal: N/A
Temperature Correction Factor (°C): 0

Departure Time: 0905

FIELD EQUIPMENT CALIBRATION

pH Calibration (SM4500-H B/9040C)				
Calibration (acceptance criteria ± 0.05 pH unit)				ICV (± 0.1 pH unit)
	7.00 Buffer	4.00 Buffer	10.00 Buffer	7.00 Buffer
	W	W	W	W
Time	0820	0822	0824	0826
Temp °C	19.2°C	19.5°C	19.7°C	19.6°C
pH	7.02	4.01	10.06	7.01

FIELD EQUIPMENT CALIBRATION

Specific Conductance (mS/cm) (99% -101%)/(mmho/cm) (SM2510 B/120.1/9050A)		
Calibration ($\pm 1\%$) (99% -101%)		ICV ($\pm 1\%$) (99% -101%)
	WP	WP
Time		
Temp °C		
Reading (mS/cm)		

Sampler Signature/Date: [Signature] 11/7/25

Supervisor Review/Date: _____

Client Name: EUNDPASTAY
Client Address: 574 JENSEY AVE
Client Rep on Site: KEVIN CARNOCCI
Sampling Date: 11-7-25
Arrival Time: 0830

Project Name: MCUB Plant No. 14245 - 5th JAMES AVE NW - 1
Project Location: New Brunswick
Cooler Custody Seal: N/A
Temperature Correction Factor (°C): 0

FIELD SAMPLING INFORMATION

[illegible]

Meter: YSI PRO Quatro, Serial # 25A103198

Sampler Signature/Date:

Supervisor Review/Date: _____

QA Control# A3041253

Page 8

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312