

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3578

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137843

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.02	7	100	90-110	11/10/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	11/10/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	11/10/2025
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	11/10/2025

Initial and Continuing Calibration Verification

Client: Europastry

SDG No.: Q3578

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

RunNo.: LB137904

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	11/14/2025
Sample ID: CCV2 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV3 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV4 COD	mg/L	49.161	50	98	95-105	11/14/2025