

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	LB137878BS
Client ID:	LB137878BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	19.0		17.8		1	6.52		11/13/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3578-01
Client ID:	MH-1172025DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	469		466		1	0.64		11/07/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3578
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: Q3578-01
Client ID: MH-1172025MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	115		113		1	1.23		11/13/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3578
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: Q3595-01
Client ID: LAW-25-0176DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.13		7.15		1	0.28		11/10/2025

Duplicate Sample Summary

Client: Europastry Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ Client ID: DSN003DUP	SDG No.: Q3578 Sample ID: Q3630-05 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	23.9		22.9		1	4.27		11/14/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3578
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: Q3630-05
Client ID: DSN003MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	68.4		67.4		1	1.47		11/14/2025