

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137848

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137848BS							
TSS	mg/L	550	531		96	1	90-110	11/11/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137871

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137871BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	11/07/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137877

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137877BS							
Oil and Grease	mg/L	20.0	18.2		91	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BS							
TPH	mg/L	20.0	19.0		95	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137878

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137878BSD							
TPH	mg/L	20.0	17.8		89	1	78-114	11/13/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3578
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137904

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137904BS							
COD	mg/L	50	53.2		106	1	90-110	11/14/2025