

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: MW-10DUP	SDG No.: Q3584 Sample ID: Q3552-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TKN	mg/L	+/-20	2.90		2.70		1	7		11/12/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3584
Project:	Edison Landfill	Sample ID:	Q3552-04
Client ID:	MW-10MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TKN	mg/L	+/-20	7.30		7.20		1	1		11/12/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3584
Project:	Edison Landfill	Sample ID:	Q3578-01
Client ID:	MH-1172025DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	469		466		1	0.64		11/07/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3584
Project:	Edison Landfill	Sample ID:	Q3578-01
Client ID:	MH-1172025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	115		113		1	1.23		11/13/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3584
Project: Edison Landfill	Sample ID: Q3584-01
Client ID: SW-1DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		11/07/2025
Cyanide	mg/L	+/-20	0.0012	U	0.0012	U	1	0		11/10/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3584
Project: Edison Landfill	Sample ID: Q3584-01
Client ID: SW-1MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.93		0.93		2	0.43		11/07/2025
Cyanide	mg/L	+/-20	0.038		0.037		1	3		11/10/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3584
Project: Edison Landfill	Sample ID: Q3584-02
Client ID: SW-2MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.60		2.60		1	0		11/07/2025
Nitrate	mg/L	+/-20	3.40		3.40		1	0		11/07/2025
Nitrite	mg/L	+/-20	3.10		3.10		1	0		11/07/2025
Bromide	mg/L	+/-20	35.2	OR	35.1	OR	1	0		11/07/2025
Sulfate	mg/L	+/-20	1370	OR	1370	OR	1	0		11/07/2025
Chloride	mg/L	+/-20	0.19	U	0.19	U	1	0		11/07/2025
Orthophosphate as P	mg/L	+/-20	4.90		5.20		1	6		11/07/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: SW-3DUP	SDG No.: Q3584 Sample ID: Q3584-03 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	12500		13000		1	3.22		11/10/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3584
Project:	Edison Landfill	Sample ID:	Q3595-01
Client ID:	LAW-25-0176DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.13		7.15		1	0.28		11/10/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3584
Project: Edison Landfill	Sample ID: Q3630-05
Client ID: DSN003DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	23.9		22.9		1	4.27		11/14/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: DSN003MSD	SDG No.: Q3584 Sample ID: Q3630-05 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	68.4		67.4		1	1.47		11/14/2025