

SOP ID : M7470A-Mercury-20

SDG No : N/A

Matrix : WATER

Pipette ID: HG A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #1

Block ID: 1. HG HOT BLOCK#3 2. N/A


Start Digest Date: 11/10/2025 **Time :** 07:40 **Temp :** 95 °C

End Digest Date: 11/10/2025 **Time :** 09:40 **Temp :** 95 °C

Digestion tube ID: M5595

Block thermometer ID: HG-DIG#3

Dig Technician Signature: 

Supervisor Signature: 

Temp : 1. 95°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP87906
CCV	30mL	MP87908
CRA	30mL	MP87910
Blank Spike	0.48mL	MP87899
Matrix Spike	0.48mL	MP87899

Chemical Used	ML/SAMPLE USED	Lot Number
HNO3/H2SO4(1:2)	2.25mL	MP87633
KMnO4 (5%)	4.5mL	MP87634
K2S2O8 (5%)	2.4mL	MP87635
Hydroxylamine HCL (12%)	1.8mL	MP87636
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP87900
0.05 ppb	S0.05	N/A	N/A
0.2 ppb	S0.2	30mL	MP87901
2.5 ppb	S2.5	30mL	MP87902
5.0 ppb	S5.0	30mL	MP87903
7.5 ppb	S7.5	30mL	MP87904
10.0 ppb	S10.0	30mL	MP87905
ICV	ICV	30mL	MP87906
ICB	ICB	30mL	MP87907
CCV	CCV	30mL	MP87908
CCB	CCB	30mL	MP87909
CRI	CRI	30mL	MP87910
CHK STD	CHK STD	30mL	MP87911

Extraction Conformance/Non-Conformance Comments:

N/A		
Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/10/2025 10:57		
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Comment	Prep Pos
PB170468BL	PBW468	30	30	<2	N/A	3-1
PB170468BS	LCS468	30	30	<2	MP87899	2
Q3561-01	1027-A-D	30	30	<2	N/A	3
Q3569-01	MW-2	30	30	<2	N/A	4
Q3569-02	MW-12	30	30	<2	N/A	5
Q3574-01	304641-01 LIQUID	30	30	<2	N/A	6
Q3584-01	SW-1	30	30	<2	N/A	7
Q3584-01DUP	SW-1DUP	30	30	<2	N/A	8
Q3584-01MS	SW-1MS	30	30	<2	MP87899	9
Q3584-01MSD	SW-1MSD	30	30	<2	MP87899	10
Q3584-02	SW-2	30	30	<2	N/A	11
Q3584-03	SW-3	30	30	<2	N/A	12
Q3584-05	FB-2	30	30	<2	N/A	13
Q3584-07	SEEP-1	30	30	<2	N/A	14