

Continuing Calibration Report for SequenceID : FF111225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016656.D

Aromatic C10-C12	5209448.000	40.000	4.507	6.384	130236.200	131963.544	1.309
Aromatic C12-C16	7860079.000	60.000	6.385	9.088	131001.317	135166.617	3.082
Aromatic C16-C21	9801270.000	80.000	9.089	13.409	122515.875	133743.127	8.395
Aromatic C21-C36	20907615.000	180.000	13.410	18.874	116153.417	117666.204	1.286
Aromatic EPH	43778412.000	360.000	4.507	18.874	121606.700	125744.182	3.290