

Continuing Calibration Report for SequenceID : FF111225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016664.D

Aromatic C10-C12	4958729.000	40.000	4.507	6.384	123968.225	131963.544	6.059
Aromatic C12-C16	7521799.000	60.000	6.385	9.088	125363.317	135166.617	7.253
Aromatic C16-C21	9402536.000	80.000	9.089	13.409	117531.700	133743.127	12.121
Aromatic C21-C36	20186096.000	180.000	13.410	18.874	112144.978	117666.204	4.692
Aromatic EPH	42069160.000	360.000	4.507	18.874	116858.778	125744.182	7.066