

Continuing Calibration Report for SequenceID : FG111325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016907.D

Aliphatic C9-C12	7691493.000	60.000	3.260	6.903	128191.550	122003.227	-5.072
Aliphatic C12-C16	5276607.000	40.000	6.904	10.362	131915.175	122851.240	-7.378
Aliphatic C16-C21	8211247.000	60.000	10.363	13.751	136854.117	132396.253	-3.367
Aliphatic C21-C28	10726527.000	80.000	13.752	17.437	134081.588	139034.659	3.562
Aliphatic C28-C40	13084899.000	120.000	17.438	22.462	109040.825	113794.545	4.177
Aliphatic EPH	44990773.000	360.000	3.260	22.462	124974.369	124878.157	-0.077