

Continuing Calibration Report for SequenceID : FG111325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016912.D

Aliphatic C9-C12	7839554.000	60.000	3.260	6.903	130659.233	122003.227	-7.095
Aliphatic C12-C16	5425659.000	40.000	6.904	10.362	135641.475	122851.240	-10.411
Aliphatic C16-C21	8437707.000	60.000	10.363	13.751	140628.450	132396.253	-6.218
Aliphatic C21-C28	10845602.000	80.000	13.752	17.437	135570.025	139034.659	2.492
Aliphatic C28-C40	13108710.000	120.000	17.438	22.462	109239.250	113794.545	4.003
Aliphatic EPH	45657232.000	360.000	3.260	22.462	126825.644	124878.157	-1.560