

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3584

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 10:45

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.64       | 6.65      | 6.55      | 6.75  | 0.01       |
| Aroclor-1254-2       | (2) | 6.86       | 6.87      | 6.77      | 6.97  | 0.01       |
| Aroclor-1254-3       | (3) | 7.22       | 7.23      | 7.13      | 7.33  | 0.01       |
| Aroclor-1254-4       | (4) | 7.50       | 7.51      | 7.41      | 7.61  | 0.01       |
| Aroclor-1254-5       | (5) | 7.92       | 7.93      | 7.83      | 8.03  | 0.01       |
| Tetrachloro-m-xylene |     | 4.64       | 4.65      | 4.55      | 4.75  | 0.01       |
| Decachlorobiphenyl   |     | 10.41      | 10.42     | 10.32     | 10.52 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3584

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 10:45

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.65       | 5.66      | 5.56      | 5.76 | 0.01       |
| Aroclor-1254-2 (2)   | 5.80       | 5.81      | 5.71      | 5.91 | 0.01       |
| Aroclor-1254-3 (3)   | 6.20       | 6.21      | 6.11      | 6.31 | 0.01       |
| Aroclor-1254-4 (4)   | 6.43       | 6.44      | 6.34      | 6.54 | 0.01       |
| Aroclor-1254-5 (5)   | 6.85       | 6.85      | 6.75      | 6.95 | 0.01       |
| Tetrachloro-m-xylene | 3.78       | 3.79      | 3.69      | 3.89 | 0.01       |
| Decachlorobiphenyl   | 8.78       | 8.79      | 8.69      | 8.89 | 0.02       |