

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3584

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 20:30

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1242-2 (2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1242-3 (3)	5.87	5.88	5.78	5.98	0.01
Aroclor-1242-4 (4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1242-5 (5)	6.70	6.71	6.61	6.81	0.01
Tetrachloro-m-xylene	4.64	4.65	4.55	4.75	0.01
Decachlorobiphenyl	10.41	10.42	10.32	10.52	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3584

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 20:30

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1242-2 (2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1242-3 (3)	5.05	5.06	4.96	5.16	0.01
Aroclor-1242-4 (4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1242-5 (5)	5.65	5.66	5.56	5.76	0.01
Tetrachloro-m-xylene	3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl	8.77	8.79	8.69	8.89	0.02