

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076435.D **Time Analyzed:** 15:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.788	5.695	5.895	574.770	500.000	15.0
Aroclor-1016-2	5.809	5.717	5.917	472.590	500.000	-5.5
Aroclor-1016-3	5.872	5.780	5.980	483.120	500.000	-3.4
Aroclor-1016-4	5.969	5.877	6.077	490.490	500.000	-1.9
Aroclor-1016-5	6.262	6.169	6.369	509.320	500.000	1.9
Aroclor-1260-1	7.381	7.287	7.487	563.820	500.000	12.8
Aroclor-1260-2	7.634	7.540	7.740	524.970	500.000	5.0
Aroclor-1260-3	7.992	7.899	8.099	543.370	500.000	8.7
Aroclor-1260-4	8.220	8.127	8.327	520.300	500.000	4.1
Aroclor-1260-5	8.545	8.452	8.652	492.730	500.000	-1.5
Decachlorobiphenyl	10.410	10.319	10.519	46.460	50.000	-7.1
Tetrachloro-m-xylene	4.635	4.542	4.742	48.360	50.000	-3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076435.D **Time Analyzed:** 15:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.866	4.778	4.978	480.580	500.000	-3.9
Aroclor-1016-2	4.925	4.836	5.036	442.930	500.000	-11.4
Aroclor-1016-3	5.045	4.954	5.154	455.510	500.000	-8.9
Aroclor-1016-4	5.086	4.996	5.196	487.960	500.000	-2.4
Aroclor-1016-5	5.300	5.210	5.410	469.110	500.000	-6.2
Aroclor-1260-1	6.328	6.238	6.438	454.930	500.000	-9.0
Aroclor-1260-2	6.515	6.425	6.625	471.600	500.000	-5.7
Aroclor-1260-3	6.668	6.578	6.778	474.380	500.000	-5.1
Aroclor-1260-4	7.137	7.047	7.247	498.500	500.000	-0.3
Aroclor-1260-5	7.378	7.287	7.487	469.890	500.000	-6.0
Decachlorobiphenyl	8.771	8.682	8.882	40.210	50.000	-19.6
Tetrachloro-m-xylene	3.773	3.681	3.881	49.430	50.000	-1.1