

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076448.D **Time Analyzed:** 19:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.695	5.895	476.690	500.000	-4.7
Aroclor-1016-2	5.813	5.717	5.917	475.640	500.000	-4.9
Aroclor-1016-3	5.875	5.780	5.980	483.470	500.000	-3.3
Aroclor-1016-4	5.973	5.877	6.077	497.820	500.000	-0.4
Aroclor-1016-5	6.265	6.169	6.369	494.470	500.000	-1.1
Aroclor-1260-1	7.385	7.287	7.487	558.440	500.000	11.7
Aroclor-1260-2	7.638	7.540	7.740	530.440	500.000	6.1
Aroclor-1260-3	7.996	7.899	8.099	542.890	500.000	8.6
Aroclor-1260-4	8.223	8.127	8.327	539.660	500.000	7.9
Aroclor-1260-5	8.549	8.452	8.652	495.520	500.000	-0.9
Decachlorobiphenyl	10.412	10.319	10.519	46.740	50.000	-6.5
Tetrachloro-m-xylene	4.640	4.542	4.742	50.750	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076448.D **Time Analyzed:** 19:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.870	4.778	4.978	490.530	500.000	-1.9
Aroclor-1016-2	4.929	4.836	5.036	466.260	500.000	-6.7
Aroclor-1016-3	5.050	4.954	5.154	442.530	500.000	-11.5
Aroclor-1016-4	5.091	4.996	5.196	485.830	500.000	-2.8
Aroclor-1016-5	5.305	5.210	5.410	448.520	500.000	-10.3
Aroclor-1260-1	6.331	6.238	6.438	518.320	500.000	3.7
Aroclor-1260-2	6.518	6.425	6.625	522.470	500.000	4.5
Aroclor-1260-3	6.671	6.578	6.778	529.570	500.000	5.9
Aroclor-1260-4	7.140	7.047	7.247	585.630	500.000	17.1
Aroclor-1260-5	7.380	7.287	7.487	512.900	500.000	2.6
Decachlorobiphenyl	8.773	8.682	8.882	41.980	50.000	-16.0
Tetrachloro-m-xylene	3.777	3.681	3.881	50.120	50.000	0.2