

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091106.D **Time Analyzed:** 14:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	51.080	50.000	2.2
4,4'-DDE	6.189	6.090	6.290	46.450	50.000	-7.1
4,4'-DDT	7.015	6.915	7.115	44.810	50.000	-10.4
Aldrin	5.263	5.164	5.364	48.170	50.000	-3.7
alpha-BHC	3.994	3.894	4.094	49.260	50.000	-1.5
alpha-Chlordane	6.020	5.921	6.121	47.660	50.000	-4.7
beta-BHC	4.512	4.412	4.612	47.370	50.000	-5.3
Decachlorobiphenyl	9.066	8.964	9.164	43.620	50.000	-12.8
delta-BHC	4.760	4.660	4.860	48.920	50.000	-2.2
Dieldrin	6.340	6.240	6.440	47.810	50.000	-4.4
Endosulfan I	6.067	5.968	6.168	47.400	50.000	-5.2
Endosulfan II	6.780	6.681	6.881	45.240	50.000	-9.5
Endosulfan sulfate	7.145	7.045	7.245	46.470	50.000	-7.1
Endrin	6.567	6.468	6.668	45.860	50.000	-8.3
Endrin aldehyde	6.910	6.810	7.010	46.530	50.000	-6.9
Endrin ketone	7.625	7.525	7.725	47.550	50.000	-4.9
gamma-BHC (Lindane)	4.324	4.225	4.425	48.390	50.000	-3.2
gamma-Chlordane	5.939	5.840	6.040	47.140	50.000	-5.7
Heptachlor	4.922	4.823	5.023	57.320	50.000	14.6
Heptachlor epoxide	5.684	5.584	5.784	48.260	50.000	-3.5
Methoxychlor	7.488	7.388	7.588	45.760	50.000	-8.5
Tetrachloro-m-xylene	3.544	3.445	3.645	47.550	50.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091106.D **Time Analyzed:** 14:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	55.810	50.000	11.6
4,4'-DDE	5.364	5.266	5.466	52.120	50.000	4.2
4,4'-DDT	6.172	6.074	6.274	50.000	50.000	0.0
Aldrin	4.358	4.261	4.461	52.430	50.000	4.9
alpha-BHC	3.385	3.287	3.487	53.050	50.000	6.1
alpha-Chlordane	5.179	5.081	5.281	51.920	50.000	3.8
beta-BHC	4.017	3.919	4.119	51.780	50.000	3.6
Decachlorobiphenyl	8.060	7.962	8.162	53.340	50.000	6.7
delta-BHC	4.253	4.155	4.355	52.470	50.000	4.9
Dieldrin	5.502	5.404	5.604	52.550	50.000	5.1
Endosulfan I	5.236	5.138	5.338	51.700	50.000	3.4
Endosulfan II	6.070	5.972	6.172	52.410	50.000	4.8
Endosulfan sulfate	6.472	6.374	6.574	52.700	50.000	5.4
Endrin	5.779	5.681	5.881	52.500	50.000	5.0
Endrin aldehyde	6.249	6.150	6.350	52.420	50.000	4.8
Endrin ketone	6.981	6.883	7.083	54.990	50.000	10.0
gamma-BHC (Lindane)	3.721	3.623	3.823	52.570	50.000	5.1
gamma-Chlordane	5.115	5.017	5.217	52.160	50.000	4.3
Heptachlor	4.073	3.975	4.175	53.730	50.000	7.5
Heptachlor epoxide	4.862	4.764	4.964	51.170	50.000	2.3
Methoxychlor	6.743	6.645	6.845	51.870	50.000	3.7
Tetrachloro-m-xylene	2.874	2.775	2.975	52.520	50.000	5.0