

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091118.D **Time Analyzed:** 17:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	51.480	50.000	3.0
4,4'-DDE	6.189	6.090	6.290	46.850	50.000	-6.3
4,4'-DDT	7.015	6.915	7.115	44.080	50.000	-11.8
Aldrin	5.263	5.164	5.364	48.180	50.000	-3.6
alpha-BHC	3.993	3.894	4.094	49.120	50.000	-1.8
alpha-Chlordane	6.020	5.921	6.121	45.510	50.000	-9.0
beta-BHC	4.511	4.412	4.612	47.330	50.000	-5.3
Decachlorobiphenyl	9.066	8.964	9.164	44.630	50.000	-10.7
delta-BHC	4.760	4.660	4.860	48.280	50.000	-3.4
Dieldrin	6.340	6.240	6.440	47.760	50.000	-4.5
Endosulfan I	6.067	5.968	6.168	47.290	50.000	-5.4
Endosulfan II	6.781	6.681	6.881	45.540	50.000	-8.9
Endosulfan sulfate	7.145	7.045	7.245	46.850	50.000	-6.3
Endrin	6.568	6.468	6.668	44.900	50.000	-10.2
Endrin aldehyde	6.910	6.810	7.010	46.630	50.000	-6.7
Endrin ketone	7.625	7.525	7.725	49.580	50.000	-0.8
gamma-BHC (Lindane)	4.324	4.225	4.425	48.270	50.000	-3.5
gamma-Chlordane	5.939	5.840	6.040	47.270	50.000	-5.5
Heptachlor	4.922	4.823	5.023	56.020	50.000	12.0
Heptachlor epoxide	5.684	5.584	5.784	47.430	50.000	-5.1
Methoxychlor	7.488	7.388	7.588	44.590	50.000	-10.8
Tetrachloro-m-xylene	3.543	3.445	3.645	47.710	50.000	-4.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091118.D **Time Analyzed:** 17:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	56.630	50.000	13.3
4,4'-DDE	5.364	5.266	5.466	52.110	50.000	4.2
4,4'-DDT	6.173	6.074	6.274	49.780	50.000	-0.4
Aldrin	4.359	4.261	4.461	51.900	50.000	3.8
alpha-BHC	3.385	3.287	3.487	52.650	50.000	5.3
alpha-Chlordane	5.180	5.081	5.281	51.770	50.000	3.5
beta-BHC	4.017	3.919	4.119	51.330	50.000	2.7
Decachlorobiphenyl	8.060	7.962	8.162	54.880	50.000	9.8
delta-BHC	4.253	4.155	4.355	51.970	50.000	3.9
Dieldrin	5.502	5.404	5.604	52.710	50.000	5.4
Endosulfan I	5.237	5.138	5.338	51.830	50.000	3.7
Endosulfan II	6.071	5.972	6.172	52.950	50.000	5.9
Endosulfan sulfate	6.472	6.374	6.574	53.350	50.000	6.7
Endrin	5.779	5.681	5.881	51.310	50.000	2.6
Endrin aldehyde	6.249	6.150	6.350	52.990	50.000	6.0
Endrin ketone	6.981	6.883	7.083	56.380	50.000	12.8
gamma-BHC (Lindane)	3.721	3.623	3.823	52.050	50.000	4.1
gamma-Chlordane	5.115	5.017	5.217	52.010	50.000	4.0
Heptachlor	4.073	3.975	4.175	52.860	50.000	5.7
Heptachlor epoxide	4.863	4.764	4.964	50.570	50.000	1.1
Methoxychlor	6.743	6.645	6.845	51.320	50.000	2.6
Tetrachloro-m-xylene	2.873	2.775	2.975	52.120	50.000	4.2