

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091130.D **Time Analyzed:** 21:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	51.780	50.000	3.6
4,4'-DDE	6.189	6.090	6.290	46.390	50.000	-7.2
4,4'-DDT	7.015	6.915	7.115	42.340	50.000	-15.3
Aldrin	5.263	5.164	5.364	47.900	50.000	-4.2
alpha-BHC	3.993	3.894	4.094	48.720	50.000	-2.6
alpha-Chlordane	6.020	5.921	6.121	45.540	50.000	-8.9
beta-BHC	4.511	4.412	4.612	46.950	50.000	-6.1
Decachlorobiphenyl	9.066	8.964	9.164	43.910	50.000	-12.2
delta-BHC	4.759	4.660	4.860	48.000	50.000	-4.0
Dieldrin	6.340	6.240	6.440	47.390	50.000	-5.2
Endosulfan I	6.067	5.968	6.168	46.890	50.000	-6.2
Endosulfan II	6.780	6.681	6.881	45.040	50.000	-9.9
Endosulfan sulfate	7.144	7.045	7.245	46.180	50.000	-7.6
Endrin	6.567	6.468	6.668	44.320	50.000	-11.4
Endrin aldehyde	6.909	6.810	7.010	46.430	50.000	-7.1
Endrin ketone	7.625	7.525	7.725	47.910	50.000	-4.2
gamma-BHC (Lindane)	4.324	4.225	4.425	47.940	50.000	-4.1
gamma-Chlordane	5.939	5.840	6.040	46.720	50.000	-6.6
Heptachlor	4.922	4.823	5.023	55.760	50.000	11.5
Heptachlor epoxide	5.683	5.584	5.784	47.360	50.000	-5.3
Methoxychlor	7.489	7.388	7.588	43.050	50.000	-13.9
Tetrachloro-m-xylene	3.544	3.445	3.645	47.140	50.000	-5.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3584
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/12/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD091130.D **Time Analyzed:** 21:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.919	5.821	6.021	56.850	50.000	13.7
4,4'-DDE	5.364	5.266	5.466	51.890	50.000	3.8
4,4'-DDT	6.172	6.074	6.274	47.530	50.000	-4.9
Aldrin	4.359	4.261	4.461	51.900	50.000	3.8
alpha-BHC	3.386	3.287	3.487	52.680	50.000	5.4
alpha-Chlordane	5.179	5.081	5.281	51.650	50.000	3.3
beta-BHC	4.018	3.919	4.119	51.280	50.000	2.6
Decachlorobiphenyl	8.059	7.962	8.162	53.920	50.000	7.8
delta-BHC	4.254	4.155	4.355	52.060	50.000	4.1
Dieldrin	5.502	5.404	5.604	52.450	50.000	4.9
Endosulfan I	5.236	5.138	5.338	51.320	50.000	2.6
Endosulfan II	6.070	5.972	6.172	52.440	50.000	4.9
Endosulfan sulfate	6.472	6.374	6.574	52.560	50.000	5.1
Endrin	5.778	5.681	5.881	50.650	50.000	1.3
Endrin aldehyde	6.248	6.150	6.350	52.610	50.000	5.2
Endrin ketone	6.981	6.883	7.083	55.390	50.000	10.8
gamma-BHC (Lindane)	3.722	3.623	3.823	52.040	50.000	4.1
gamma-Chlordane	5.114	5.017	5.217	51.750	50.000	3.5
Heptachlor	4.073	3.975	4.175	52.980	50.000	6.0
Heptachlor epoxide	4.862	4.764	4.964	50.700	50.000	1.4
Methoxychlor	6.743	6.645	6.845	49.480	50.000	-1.0
Tetrachloro-m-xylene	2.874	2.775	2.975	51.990	50.000	4.0