

Surrogate Summary

SDG No.: **Q3584**

Client: **Remington & Vernick Engineers**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.4	97		30 (41)	150 (134)
		Decachlorobiphen	2	20	20.3	102		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	20.0	100		30 (41)	150 (134)
I.BLK-PD091081.D	PIBLK-PD091081.D	Decachlorobiphen	1	20	21.9	109		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	21.8	109		30 (41)	150 (134)
		Decachlorobiphen	2	20	27.4	137		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	25.6	128		30 (41)	150 (134)
PB170485BL	PB170485BL	Decachlorobiphen	1	20	20.7	103		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.4	102		30 (41)	150 (134)
		Decachlorobiphen	2	20	25.6	128		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	23.6	118		30 (41)	150 (134)
PB170485BS	PB170485BS	Decachlorobiphen	1	20	21.4	107		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.5	103		30 (41)	150 (134)
		Decachlorobiphen	2	20	26.4	132		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	23.9	120		30 (41)	150 (134)
PB170485BSD	PB170485BSD	Decachlorobiphen	1	20	17.5	87		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.9	99		30 (41)	150 (134)
		Decachlorobiphen	2	20	20.7	104		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	23.1	115		30 (41)	150 (134)
Q3584-01	SW-1	Decachlorobiphen	1	20	18.0	90		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.0	100		30 (41)	150 (134)
		Decachlorobiphen	2	20	19.9	99		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	22.9	115		30 (41)	150 (134)
Q3584-02	SW-2	Decachlorobiphen	1	20	20.6	103		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.8	104		30 (41)	150 (134)
		Decachlorobiphen	2	20	21.9	109		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	23.8	119		30 (41)	150 (134)
Q3584-03	SW-3	Decachlorobiphen	1	20	19.4	97		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	21.0	105		30 (41)	150 (134)
		Decachlorobiphen	2	20	21.1	106		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	24.3	121		30 (41)	150 (134)
Q3584-04	EB-2	Decachlorobiphen	1	20	14.4	72		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.3	102		30 (41)	150 (134)
		Decachlorobiphen	2	20	15.1	75		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	22.8	114		30 (41)	150 (134)
Q3584-05	FB-2	Decachlorobiphen	1	20	17.0	85		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.2	101		30 (41)	150 (134)
		Decachlorobiphen	2	20	20.3	102		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	22.5	113		30 (41)	150 (134)
I.BLK-PD091104.D	PIBLK-PD091104.D	Decachlorobiphen	1	20	21.7	108		30 (31)	150 (149)

Surrogate Summary

SDG No.: Q3584

Client: Remington & Vernick Engineers

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD091104.D	PIBLK-PD091104.D	Tetrachloro-m-xyl	1	20	22.0	110		30 (41)	150 (134)
		Decachlorobiphen	2	20	25.9	130		30 (31)	150 (149)
I.BLK-PD091117.D	PIBLK-PD091117.D	Tetrachloro-m-xyl	2	20	24.7	124		30 (41)	150 (134)
		Decachlorobiphen	1	20	21.7	108		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	21.5	107		30 (41)	150 (134)
		Decachlorobiphen	2	20	25.9	129		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	24.3	121		30 (41)	150 (134)
Q3584-07	SEEP-1	Decachlorobiphen	1	20	9.43	47		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	11.7	58		30 (41)	150 (134)
		Decachlorobiphen	2	20	7.61	38		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	13.1	66		30 (41)	150 (134)
I.BLK-PD091129.D	PIBLK-PD091129.D	Decachlorobiphen	1	20	22.1	110		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	21.6	108		30 (41)	150 (134)
		Decachlorobiphen	2	20	26.1	131		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	24.5	123		30 (41)	150 (134)