

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3584</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>11/12/2025 09:31</u>
Lab File ID:	<u>BN038183.D</u>	Init. Calib. Date(s):	<u>10/28/2025 10/28/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>13:00 16:38</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.571	0.558		-2.3	20.0
Fluoranthene-d10	1.008	0.917		-9.0	20.0
2-Fluorophenol	0.943	0.939		-0.4	20.0
Phenol-d6	1.148	1.155		0.6	20.0
Nitrobenzene-d5	0.323	0.321		-0.6	20.0
2-Fluorobiphenyl	1.602	1.475		-7.9	20.0
2,4,6-Tribromophenol	0.165	0.138		-16.4	20.0
Terphenyl-d14	0.869	0.919		5.8	20.0
1,4-Dioxane	0.416	0.395		-5.0	20.0

All other compounds must meet a minimum RRF of 0.010.