

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111325\
 Data File : BN038198.D
 Acq On : 13 Nov 2025 12:14
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 13 12:55:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.377	5.8	101	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	108	0.00
4 S	2-Fluorophenol	0.400	0.400	0.0	108	0.00
5 S	Phenol-d6	0.400	0.413	-3.2	114	0.00
6	bis(2-Chloroethyl)ether	0.400	0.393	1.8	107	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	118	-0.01
9	Naphthalene	0.400	0.395	1.3	115	-0.01
10	Hexachlorobutadiene	0.400	0.399	0.3	113	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.396	1.0	117	0.00
12	2-Methylnaphthalene	0.400	0.396	1.0	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.338	15.5	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.367	8.3	112	0.00
16	Acenaphthylene	0.400	0.375	6.3	119	0.00
17	Acenaphthene	0.400	0.376	6.0	118	0.00
18	Fluorene	0.400	0.376	6.0	117	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.371	7.3	92	0.00
21	4-Bromophenyl-phenylether	0.400	0.391	2.3	119	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	120	-0.01
23	Atrazine	0.400	0.347	13.3	111	0.00
24	Pentachlorophenol	0.400	0.264	34.0#	92	0.00
25	Phenanthrene	0.400	0.385	3.8	116	0.00
26	Anthracene	0.400	0.370	7.5	114	-0.01
27 SURR	Fluoranthene-d10	0.400	0.355	11.3	109	0.00
28	Fluoranthene	0.400	0.360	10.0	111	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	101	0.00
30	Pyrene	0.400	0.438	-9.5	109	0.00
31 S	Terphenyl-d14	0.400	0.443	-10.7	112	0.00
32	Benzo(a)anthracene	0.400	0.378	5.5	101	0.00
33	Chrysene	0.400	0.394	1.5	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.472	-18.0	125	0.00
35 I	Perylene-d12	0.400	0.400	0.0	106	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.433	-8.2	117	0.00
37	Benzo(b)fluoranthene	0.400	0.377	5.8	104	0.00
38	Benzo(k)fluoranthene	0.400	0.385	3.8	108	0.00
39 C	Benzo(a)pyrene	0.400	0.394	1.5	108	0.00
40	Dibenzo(a,h)anthracene	0.400	0.430	-7.5	120	0.00
41	Benzo(g,h,i)perylene	0.400	0.428	-7.0	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4