



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111325\
Data File : BN038197.D
Acq On : 13 Nov 2025 11:34
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 29 13:09:46 2025






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.0	2754	PASS
69	69	100	100	100.0	266965	PASS
70	69	0.00	2	0.5	1417	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	715200	PASS
199	198	5	9	6.9	49421	PASS
365	198	1	100	3.4	24224	PASS
441	443	0.01	150	85.3	103765	PASS
442	442	100	100	100.0	623872	PASS
443	442	15	24	19.5	121616	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111325\
Data File : BN038197.D
Acq On : 13 Nov 2025 11:34
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 13 17:42:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 13 17:42:14 2025
Response via : Initial Calibration



TIC: BN038197.D\data.ms

(70) Pentachlorophenol (C)

16.845min (0.000) 277.89 ng




response 1078097

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.33
264.00	61.60	62.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111325\
Data File : BN038197.D
Acq On : 13 Nov 2025 11:34
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Nov 13 17:42:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 13 17:42:14 2025
Response via : Initial Calibration



TIC: BN038197.D\data.ms

(77) Benzidine

19.445min (0.000) 0.00 ng

response 4694755

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.03
183.00	13.20	12.05
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/13/2025	BNA_N	<u>BN038197.D</u>
Compound Name	Response	Retention Time
DDT	2779111	20.68
DDD	50940	20.239
DDE	9135	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
60075	2839186	2.12

Instrument :
BNA_N
ClientSampleId :
DFTPP