

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038110.D	28 Oct 2025 12:05	RC/JU	Ok
2	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00	RC/JU	Ok
3	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36	RC/JU	Ok
4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	RC/JU	Ok
5	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49	RC/JU	Ok,M
6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25	RC/JU	Ok
7	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02	RC/JU	Ok
8	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	RC/JU	Ok
9	SSTDICV0.4	BN038118.D	28 Oct 2025 17:51	RC/JU	Ok
10	PB170167BL	BN038119.D	28 Oct 2025 19:04	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN111225

Review By	rahul	Review On	11/12/2025 1:44:37 PM
Supervise By	Jagrut	Supervise On	11/13/2025 10:06:02 AM
SubDirectory	BN111225	HP Acquire Method	BNA_N
		HP Processing Method	BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038182.D	12 Nov 2025 08:52	RC/JU	Ok
2	SSTDCCC0.4	BN038183.D	12 Nov 2025 09:31	RC/JU	Ok
3	PB170483BL	BN038184.D	12 Nov 2025 10:07	RC/JU	Ok
4	PB170483BS	BN038185.D	12 Nov 2025 10:44	RC/JU	Ok
5	PB170483BSD	BN038186.D	12 Nov 2025 11:20	RC/JU	Ok
6	Q3569-02	BN038187.D	12 Nov 2025 11:56	RC/JU	Dilution
7	Q3584-01	BN038188.D	12 Nov 2025 12:32	RC/JU	Ok
8	Q3569-01	BN038189.D	12 Nov 2025 13:09	RC/JU	Dilution
9	Q3584-03	BN038190.D	12 Nov 2025 13:45	RC/JU	Ok
10	Q3584-02	BN038191.D	12 Nov 2025 14:21	RC/JU	Ok
11	Q3584-04	BN038192.D	12 Nov 2025 14:58	RC/JU	Ok
12	Q3584-05	BN038193.D	12 Nov 2025 15:34	RC/JU	Ok
13	Q3569-02DL	BN038194.D	12 Nov 2025 16:11	RC/JU	Ok
14	Q3569-01DL	BN038195.D	12 Nov 2025 16:47	RC/JU	Ok
15	SSTDCCC0.4	BN038196.D	12 Nov 2025 18:40	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN111325

Review By	Jagrut	Review On	11/13/2025 5:04:18 PM
Supervise By	Sohil	Supervise On	11/14/2025 12:27:41 PM
SubDirectory	BN111325	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN102825
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904		
CCC	SP6902		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6906		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038197.D	13 Nov 2025 11:34	RC/JU	Ok
2	SSTDCCC0.4	BN038198.D	13 Nov 2025 12:14	RC/JU	Ok
3	PB170483BL	BN038199.D	13 Nov 2025 12:50	RC/JU	Not Ok
4	Q3584-07	BN038200.D	13 Nov 2025 13:27	RC/JU	Dilution
5	Q3584-07DL	BN038201.D	13 Nov 2025 15:52	RC/JU	Ok
6	SSTDCCC0.4	BN038202.D	13 Nov 2025 17:07	RC/JU	Ok

M : Manual Integration