

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3584

Client ID : SSTDCCC0.4

Date Analyzed: 11/13/2025

Lab File ID: BN038198.D

Time Analyzed: 12:14

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	10040	7.768	31592	10.57	18577	14.43
UPPER LIMIT	20080	8.268	63184	11.073	37154	14.93
LOWER LIMIT	5020	7.268	15796	10.073	9288.5	13.93
EPA SAMPLE NO.						
01 SEEP-1	538910 *	7.77	1785290 *	10.57	1021410 *	14.43
02 SEEP-1DL	110160 *	7.78	385141 *	10.58	236632 *	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.