

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038110.D
Instrument ID: BNA_N

Contract: REMI02
SDG NO.: Q3584
DFTPP Injection Date: 10/28/2025
DFTPP Injection Time: 12:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	38.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	14.2
442	Greater than 50% of mass 198	85.3
443	15.0 - 24.0% of mass 442	17 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN038111.D	10/28/2025	13:00
SSTDICC0.2	SSTDICC0.2	BN038112.D	10/28/2025	13:36
SSTDICCC0.4	SSTDICCC0.4	BN038113.D	10/28/2025	14:13
SSTDICC0.8	SSTDICC0.8	BN038114.D	10/28/2025	14:49
SSTDICC1.6	SSTDICC1.6	BN038115.D	10/28/2025	15:25
SSTDICC3.2	SSTDICC3.2	BN038116.D	10/28/2025	16:02
SSTDICC5.0	SSTDICC5.0	BN038117.D	10/28/2025	16:38