

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038182.D
Instrument ID: BNA_N

Contract: REMI02
SDG NO.: Q3584
DFTPP Injection Date: 11/12/2025
DFTPP Injection Time: 08:52

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	13.1
442	Greater than 50% of mass 198	84.1
443	15.0 - 24.0% of mass 442	15.7 (18.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038183.D	11/12/2025	09:31
PB170483BL	PB170483BL	BN038184.D	11/12/2025	10:07
PB170483BS	PB170483BS	BN038185.D	11/12/2025	10:44
PB170483BSD	PB170483BSD	BN038186.D	11/12/2025	11:20
SW-1	Q3584-01	BN038188.D	11/12/2025	12:32
SW-3	Q3584-03	BN038190.D	11/12/2025	13:45
SW-2	Q3584-02	BN038191.D	11/12/2025	14:21
EB-2	Q3584-04	BN038192.D	11/12/2025	14:58
FB-2	Q3584-05	BN038193.D	11/12/2025	15:34