

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3584</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>11/11/2025 14:44</u>
Lab File ID:	<u>BP026090.D</u>	Init. Calib. Date(s):	<u>10/29/2025 10/29/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>09:26 14:13</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.212	1.264		4.3	
Benzaldehyde	0.802	0.731		-8.9	
Phenol-d6	1.543	1.614		4.6	
Phenol	1.713	1.746		1.9	20.0
bis(2-Chloroethyl)ether	1.352	1.365		1.0	
2-Chlorophenol	1.342	1.449		8.0	
2-Methylphenol	1.121	1.171		4.5	
2,2-oxybis(1-Chloropropane)	2.043	1.910		-6.5	
Acetophenone	0.506	0.507		0.2	
3+4-Methylphenols	1.560	1.626		4.2	
n-Nitroso-di-n-propylamine	0.958	0.999	0.050	4.3	
Nitrobenzene-d5	0.321	0.350		9.0	
Hexachloroethane	0.533	0.569		6.8	
Nitrobenzene	0.339	0.359		5.9	
Isophorone	0.687	0.710		3.3	
2-Nitrophenol	0.104	0.185		77.9	20.0
2,4-Dimethylphenol	0.289	0.300		3.8	
bis(2-Chloroethoxy)methane	0.433	0.445		2.8	
2,4-Dichlorophenol	0.305	0.340		11.5	20.0
Naphthalene	1.091	1.100		0.8	
4-Chloroaniline	0.419	0.430		2.6	
Hexachlorobutadiene	0.211	0.223		5.7	20.0
Caprolactam	0.107	0.117		9.3	
4-Chloro-3-methylphenol	0.320	0.353		10.3	20.0
2-Methylnaphthalene	0.763	0.781		2.4	
Hexachlorocyclopentadiene	0.227	0.287	0.050	26.4	
2,4,6-Trichlorophenol	0.367	0.434		18.3	20.0
2-Fluorobiphenyl	1.392	1.380		-0.9	
2,4,5-Trichlorophenol	0.419	0.475		13.4	
1,1-Biphenyl	1.531	1.524		-0.5	
2-Chloronaphthalene	1.205	1.209		0.3	
2-Nitroaniline	0.272	0.334		22.8	
Dimethylphthalate	1.455	1.529		5.1	
Acenaphthylene	1.756	1.763		0.4	
2,6-Dinitrotoluene	0.237	0.310		30.8	
3-Nitroaniline	0.291	0.359		23.4	
Acenaphthene	1.206	1.180		-2.2	20.0
2,4-Dinitrophenol	0.100	0.159	0.050	59.0	
4-Nitrophenol	0.290	0.317	0.050	9.3	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3584</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>11/11/2025 14:44</u>
Lab File ID:	<u>BP026090.D</u>	Init. Calib. Date(s):	<u>10/29/2025 10/29/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>09:26 14:13</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.822	1.817		-0.3	
2,4-Dinitrotoluene	0.347	0.459		32.3	
Diethylphthalate	1.463	1.536		5.0	
4-Chlorophenyl-phenylether	0.713	0.734		2.9	
Fluorene	1.427	1.432		0.3	
4-Nitroaniline	0.311	0.370		19.0	
4,6-Dinitro-2-methylphenol	0.077	0.116		50.6	
n-Nitrosodiphenylamine	0.625	0.623		-0.3	20.0
2,4,6-Tribromophenol	0.216	0.265		22.7	
4-Bromophenyl-phenylether	0.222	0.233		5.0	
Hexachlorobenzene	0.253	0.264		4.3	
Atrazine	0.211	0.223		5.7	
Pentachlorophenol	0.154	0.182		18.2	20.0
Phenanthrene	1.164	1.133		-2.7	
Anthracene	1.154	1.149		-0.4	
Carbazole	1.092	1.094		0.2	
Di-n-butylphthalate	1.222	1.365		11.7	
Fluoranthene	1.358	1.371		1.0	20.0
Pyrene	1.353	1.374		1.6	
Terphenyl-d14	0.984	0.992		0.8	
Butylbenzylphthalate	0.432	0.610		41.2	
3,3-Dichlorobenzidine	0.443	0.491		10.8	
Benzo (a) anthracene	1.374	1.392		1.3	
Chrysene	1.302	1.303		0.1	
Bis(2-ethylhexyl)phthalate	0.718	0.915		27.4	
Di-n-octyl phthalate	1.191	1.622		36.2	20.0
Benzo (b) fluoranthene	1.242	1.261		1.5	
Benzo (k) fluoranthene	1.268	1.232		-2.8	
Benzo (a) pyrene	1.141	1.172		2.7	20.0
Indeno (1,2,3-cd) pyrene	1.476	1.518		2.8	
Dibenzo (a,h) anthracene	1.202	1.234		2.7	
Benzo (g,h,i) perylene	1.156	1.176		1.7	
1,2,4,5-Tetrachlorobenzene	0.621	0.638		2.7	
1,4-Dioxane	0.511	0.508		-0.6	20.0
2,3,4,6-Tetrachlorophenol	0.330	0.408		23.6	

All other compounds must meet a minimum RRF of 0.010.