

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3584

Client ID : SSTDCCC040

Date Analyzed: 11/11/2025

Lab File ID: BF144198.D

Time Analyzed: 10:09

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	73404	6.875	271454	8.16	150816	9.92
UPPER LIMIT	146808	7.375	542908	8.657	301632	10.416
LOWER LIMIT	36702	6.375	135727	7.657	75408	9.416
EPA SAMPLE NO.						
01 FB-2	61823	6.88	239622	8.16	129099	9.92
02 SW-2	59255	6.88	231451	8.16	119771	9.92
03 EB-2	58254	6.88	216777	8.16	114918	9.92
04 SW-3	54285	6.88	195487	8.16	102900	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.